

Test report - PMFT 1000 - Oil penetration



Operator	Date	Time
Sager	13.04.2021	16:15:24

Basic information

Operation parameter	
Script	Efficiency Test (oil)
Aerosol	Oil
Main air	94.97 l/min
Generator air	2.50 l/min

Test conditions	
Mass concentration	13.95 mg/m ³


Mask information

Mask ID	13214_20210413_161524
Mask los	2060 6310223
Mask type	FFP2 / x / x

Result

Penetration	Measurement	Upper limit	Result
P_SARS_CoV-2 (0.15 µm)	0.24 %		
P_m	0.18 %		
P_Ph_45°	0.12 %	6.00 %	✓ pass

Distributions



Calibration valid until 17.08.2021